

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.8uS / 0.5ppm  
CYCLE NUMBER 001435  
12:51 09/15/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:29  
PRESSURIZING 6:08  
STERILIZING 7:40  
136.0 C 220kPa 7:40  
136.4 C 222kPa 8:10  
136.2 C 221kPa 8:40  
136.1 C 218kPa 9:10  
136.3 C 222kPa 9:40  
136.4 C 223kPa 10:10  
136.4 C 221kPa 10:40  
136.4 C 223kPa 11:10  
VENTING 11:10  
AIR DRYING 12:12  
CYCLE COMPLETE 21:16  
-----

Digital Signature #  
C9D0610828B2D2C4  
-----