
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002657
12:54 07/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:14
136.4 C 226kPa 4:14
136.3 C 225kPa 4:44
136.4 C 224kPa 5:14
136.3 C 224kPa 5:44
136.4 C 225kPa 6:14
136.4 C 224kPa 6:44
136.3 C 225kPa 7:14
136.3 C 225kPa 7:44
136.3 C 225kPa 7:44
VENTING 7:44
AIR DRYING 8:38
CYCLE COMPLETE 9:13

Digital Signature #
F67A66435C37F3C7
