
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004060
8:39 20/07/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:02
STERILIZING 3:05
136.4 C 221kPa 3:05
136.4 C 223kPa 3:35
136.1 C 217kPa 4:05
136.4 C 220kPa 4:35
136.4 C 220kPa 5:05
136.2 C 222kPa 5:35
136.3 C 220kPa 6:05
136.3 C 223kPa 6:35
136.3 C 223kPa 6:35
VENTING 6:35
AIR DRYING 8:18
CYCLE COMPLETE 8:41

Digital Signature #
2B419213A9631837
