
Midmark M11 - v1.0.5
Total Cycles: 3842

Name: RITTER2
Sterilizer ID: V2313228
08/08/2024 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.0 F 0.0
2:00 141.4 F 2.1
4:00 152.3 F 5.0
6:00 182.5 F 8.8
8:00 224.8 F 13.3
10:00 223.1 F 7.5
12:00 231.9 F 12.2
14:00 232.7 F 8.8
16:00 239.7 F 11.1
18:00 244.6 F 12.9
20:00 255.8 F 18.9
22:00 266.1 F 25.4
24:00 270.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 271.6 F 29.6
2:00 271.5 F 29.5
3:00 272.2 F 29.6
4:00 272.4 F 29.8

Min 271.2 F 29.5
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 241.0 F 0.0
10:00 225.6 F 0.0
15:00 226.9 F 0.0
20:00 210.5 F 0.0
25:00 184.4 F 0.0
30:00 180.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:03:22
CYCLE COMPLETE

08/08/2024 03:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____