

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2937

Name: RITTER2  
Sterilizer ID: V2313228  
09/27/2023 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.7 F 0.0  
2:00 171.9 F 3.1  
4:00 195.2 F 8.0  
6:00 227.0 F 10.0  
8:00 232.1 F 9.6  
10:00 234.5 F 8.6  
12:00 245.2 F 13.2  
14:00 244.0 F 12.4  
16:00 257.2 F 19.3  
18:00 268.9 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.7 F 29.5  
2:00 272.8 F 29.5  
3:00 272.6 F 29.6  
4:00 272.1 F 29.5  
  
Min 270.9 F 28.4  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
5:00 247.4 F 0.0  
10:00 205.6 F 0.0  
15:00 200.3 F 0.0  
20:00 192.4 F 0.0  
25:00 191.5 F 0.0  
30:00 188.0 F 0.0  
  
Drying Complete

FILLING: 1:52  
HEATING: 18:57  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

09/27/2023 12:01 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_