

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4054

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2024 08:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 93.7 F 0.0  
4:00 130.7 F 0.0  
6:00 149.8 F 0.1  
8:00 176.9 F 0.3  
10:00 171.8 F 1.2  
12:00 206.7 F 4.6  
14:00 207.4 F 8.5  
16:00 219.8 F 12.5  
18:00 229.7 F 7.2  
20:00 241.2 F 11.8  
22:00 236.8 F 9.2  
24:00 239.0 F 9.8  
26:00 247.7 F 14.2  
28:00 259.4 F 20.7  
30:00 270.0 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.6 F 29.4  
2:00 271.6 F 28.6  
3:00 272.6 F 29.5  
4:00 272.1 F 28.7  
  
Min 271.4 F 28.5  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 235.3 F 0.0  
10:00 184.6 F 0.0  
15:00 176.2 F 0.0  
20:00 173.6 F 0.0  
25:00 172.9 F 0.0  
30:00 173.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:55  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:09:05  
CYCLE COMPLETE

10/25/2024 09:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_