
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003045
15:46 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:15
136.0 C 220kPa 4:15
136.0 C 223kPa 4:45
136.3 C 222kPa 5:15
136.3 C 224kPa 5:45
136.3 C 223kPa 6:15
136.4 C 225kPa 6:45
136.4 C 223kPa 7:15
136.4 C 224kPa 7:45
136.4 C 224kPa 7:45
VENTING 7:45
AIR DRYING 8:40
CYCLE COMPLETE 8:42

Digital Signature #
CB0F7C65DA9CE3A8
