

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000670  
10:58 14/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:05  
STERILIZING 4:11  
136.0 C 229kPa 4:11  
136.4 C 233kPa 4:41  
136.3 C 232kPa 5:11  
136.3 C 232kPa 5:41  
136.3 C 232kPa 6:11  
136.2 C 230kPa 6:41  
136.4 C 233kPa 7:11  
136.3 C 232kPa 7:41  
136.3 C 232kPa 7:41  
VENTING 7:41  
AIR DRYING 8:36  
CYCLE COMPLETE 8:38  
-----

Digital Signature #  
5405B758CDE831C3  
-----