
Midmark M11 - v1.0.5
Total Cycles: 3072

Name: RITTER2
Sterilizer ID: V2313228
11/01/2023 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.8 F 0.0
2:00 130.8 F 0.7
4:00 155.3 F 3.4
6:00 177.3 F 7.1
8:00 220.3 F 11.5
10:00 222.8 F 5.2
12:00 233.8 F 9.7
14:00 234.7 F 9.4
16:00 240.4 F 11.9
18:00 239.7 F 10.5
20:00 252.4 F 16.9
22:00 264.2 F 23.8
24:00 271.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 273.1 F 29.9
2:00 271.6 F 29.3
3:00 272.9 F 29.6
4:00 271.7 F 29.5

Min 271.0 F 28.7
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 234.8 F 0.0
10:00 223.3 F 0.0
15:00 188.0 F 0.0
20:00 184.7 F 0.0
25:00 181.5 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

11/01/2023 03:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____