

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3554

Name: RITTER 1  
Sterilizer ID: V2178198  
07/20/2023 08:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 214.9 F 0.0  
2:00 205.0 F 1.6  
4:00 204.8 F 5.4  
6:00 214.7 F 10.6  
8:00 222.6 F 5.4  
10:00 234.6 F 10.9  
12:00 233.8 F 8.2  
14:00 241.0 F 10.6  
16:00 246.7 F 13.5  
18:00 259.2 F 20.3  
20:00 270.3 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 29.0  
2:00 271.9 F 28.4  
3:00 271.5 F 28.5  
4:00 272.6 F 29.0  
  
Min 271.3 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 218.1 F 0.0  
10:00 233.4 F 0.0  
15:00 229.3 F 0.0  
20:00 227.2 F 0.0  
25:00 224.6 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 20:55  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

07/20/2023 09:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_