
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001323
15:11 02/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:01
STERILIZING 4:00
136.0 C 220kPa 4:00
136.3 C 224kPa 4:30
136.2 C 221kPa 5:00
136.4 C 222kPa 5:30
136.4 C 223kPa 6:00
136.3 C 223kPa 6:30
136.3 C 224kPa 7:00
136.3 C 224kPa 7:30
136.3 C 224kPa 7:30
VENTING 7:30
AIR DRYING 8:27
CYCLE COMPLETE 8:39

Digital Signature #
06029C0FF33C3422
