

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1167

Name: RITTER 1  
Sterilizer ID: V2178198  
09/14/2021 08:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.8 F 0.0  
2:00 146.7 F 0.7  
4:00 171.3 F 3.1  
6:00 190.4 F 6.8  
8:00 207.5 F 11.7  
10:00 215.9 F 5.5  
12:00 232.2 F 10.2  
14:00 235.2 F 7.8  
16:00 242.3 F 12.5  
18:00 241.0 F 10.9  
20:00 254.0 F 17.5  
22:00 265.6 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.5  
1:00 272.7 F 29.2  
2:00 272.0 F 28.5  
3:00 272.7 F 29.2  
4:00 272.4 F 28.7

Min 271.7 F 28.3  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 197.3 F 0.0  
10:00 213.4 F 0.0  
15:00 215.1 F 0.0  
20:00 214.9 F 0.0  
25:00 195.8 F 0.0  
30:00 193.0 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 23:34  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:48  
CYCLE COMPLETE

09/14/2021 09:14 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01