
Midmark M11 - v1.0.0
Total Cycles: 2046

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2022 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 173.6 F 2.6
4:00 192.1 F 6.0
6:00 211.9 F 10.8
8:00 218.0 F 4.4
10:00 231.4 F 9.4
12:00 236.9 F 10.0
14:00 240.5 F 11.3
16:00 237.5 F 9.3
18:00 251.1 F 15.9
20:00 262.5 F 22.4
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.2
2:00 271.9 F 28.4
3:00 272.1 F 29.1
4:00 272.4 F 28.9

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 223.0 F 0.0
10:00 231.4 F 0.0
15:00 225.2 F 0.0
20:00 228.5 F 0.0
25:00 230.2 F 0.0
30:00 236.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

06/22/2022 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____