
Midmark M11 - v1.0.0
Total Cycles: 1976

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2022 04:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 192.2 F 4.1
4:00 206.2 F 9.5
6:00 219.6 F 6.3
8:00 232.8 F 10.1
10:00 234.3 F 8.2
12:00 244.9 F 12.6
14:00 245.5 F 12.8
16:00 258.5 F 19.8
18:00 269.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 28.0
3:00 272.4 F 28.9
4:00 272.6 F 28.6

Min 271.3 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 209.5 F 0.0
10:00 219.5 F 0.0
15:00 222.3 F 0.0
20:00 222.0 F 0.0
25:00 222.8 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

05/31/2022 06:56 AM
ACCEPTED: X
REJECTED:
Operator: U02