

-----  
Midmark M11 - v1.0.5  
Total Cycles: 344

Name: RITTER2  
Sterilizer ID: V2313228  
09/09/2021 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.8 F 0.0  
2:00 153.7 F 1.9  
4:00 164.1 F 5.3  
6:00 189.6 F 9.9  
8:00 206.9 F 5.9  
10:00 227.6 F 9.9  
12:00 231.1 F 8.4  
14:00 241.9 F 12.9  
16:00 241.7 F 11.7  
18:00 254.2 F 18.1  
20:00 265.2 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 271.6 F 29.2  
2:00 271.9 F 29.4  
3:00 271.9 F 29.4  
4:00 271.6 F 29.2

Min 271.5 F 28.9  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 230.2 F 0.0  
10:00 220.5 F 0.0  
15:00 220.6 F 0.0  
20:00 220.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:53  
HEATING: 21:41  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 20:19  
TOTAL CYCLE: 00:50:56

09/09/2021 09:45 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_