

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3552

Name: RITTER 1  
Sterilizer ID: V2178198  
07/19/2023 02:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.3 F 0.0  
2:00 191.5 F 3.8  
4:00 208.4 F 8.4  
6:00 230.3 F 10.7  
8:00 232.5 F 9.3  
10:00 239.1 F 9.2  
12:00 243.8 F 12.7  
14:00 242.8 F 11.6  
16:00 256.0 F 18.4  
18:00 267.7 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.5 F 28.6  
2:00 271.8 F 28.6  
3:00 272.1 F 28.5  
4:00 271.7 F 28.6  
  
Min 271.2 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 228.1 F 0.0  
10:00 237.4 F 0.0  
15:00 232.4 F 0.0  
20:00 228.5 F 0.0  
25:00 220.0 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 19:17  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:29  
CYCLE COMPLETE

07/19/2023 03:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_