
Midmark M11 - v1.0.5
Total Cycles: 2741

Name: RITTER2
Sterilizer ID: V2313228
08/09/2023 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 76.9 F 0.0
4:00 75.2 F 0.0
6:00 98.5 F 0.0
8:00 133.0 F 0.0
10:00 163.0 F 0.1
12:00 183.7 F 1.9
14:00 207.2 F 4.8
16:00 212.8 F 8.2
18:00 224.0 F 11.6
20:00 224.5 F 6.0
22:00 234.1 F 10.8
24:00 234.2 F 8.1
26:00 243.7 F 13.0
28:00 241.4 F 11.4
30:00 253.4 F 17.5
32:00 263.8 F 23.8
34:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.6
2:00 271.6 F 28.7
3:00 272.2 F 29.7
4:00 272.0 F 29.1

Min 271.3 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 220.0 F 0.0
10:00 199.1 F 0.0
15:00 184.0 F 0.0
20:00 200.6 F 0.0
25:00 205.0 F 0.0
30:00 197.8 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 34:12
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:12:13
CYCLE COMPLETE

08/09/2023 10:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____