
Midmark M11 - v1.0.5
Total Cycles: 2666

Name: RITTER2
Sterilizer ID: V2313228
07/19/2023 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.3 F 0.0
2:00 138.0 F 0.0
4:00 144.9 F 0.3
6:00 178.4 F 4.1
8:00 203.1 F 8.6
10:00 217.2 F 13.3
12:00 225.6 F 8.0
14:00 234.9 F 12.9
16:00 237.1 F 10.2
18:00 236.5 F 8.7
20:00 248.6 F 14.9
22:00 260.0 F 21.2
24:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.0 F 29.6
2:00 273.3 F 29.9
3:00 272.0 F 29.2
4:00 272.4 F 29.9

Min 271.0 F 28.7
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 216.7 F 0.0
10:00 192.6 F 0.0
15:00 191.7 F 0.0
20:00 190.9 F 0.0
25:00 191.1 F 0.0
30:00 203.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

07/19/2023 10:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____