
Midmark M11 - v1.0.0
Total Cycles: 1243

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2021 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.8 F 0.0
2:00 174.8 F 3.8
4:00 192.1 F 8.4
6:00 217.6 F 10.9
8:00 227.9 F 8.9
10:00 237.9 F 10.0
12:00 242.7 F 12.2
14:00 241.3 F 10.9
16:00 255.3 F 17.9
18:00 267.2 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.2 F 28.5
2:00 271.1 F 28.2
3:00 272.4 F 28.6
4:00 271.2 F 27.9

Min 271.0 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 233.7 F 0.0
10:00 211.9 F 0.0
15:00 219.9 F 0.0
20:00 232.6 F 0.0
25:00 228.3 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

10/07/2021 03:26 PM
ACCEPTED: X
REJECTED:
Operator: U02