
Midmark M11 - v1.0.0
Total Cycles: 1860

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 08:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 161.9 F 0.9
4:00 169.2 F 4.0
6:00 190.9 F 8.3
8:00 211.4 F 13.4
10:00 224.1 F 7.1
12:00 236.5 F 12.2
14:00 236.0 F 9.2
16:00 239.2 F 9.4
18:00 248.1 F 14.1
20:00 260.1 F 20.9
22:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.3 F 28.6
2:00 272.4 F 29.1
3:00 271.6 F 28.2
4:00 272.6 F 28.9

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 217.7 F 0.0
10:00 217.5 F 0.0
15:00 214.0 F 0.0
20:00 220.2 F 0.0
25:00 233.1 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

04/25/2022 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____