
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007629
9:45 03/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:03
STERILIZING 3:02
136.0 C 220kPa 3:02
136.1 C 220kPa 3:32
136.3 C 223kPa 4:02
136.3 C 224kPa 4:32
136.4 C 225kPa 5:02
136.4 C 220kPa 5:32
136.4 C 222kPa 6:02
136.4 C 224kPa 6:32
136.4 C 224kPa 6:32
VENTING 6:32
AIR DRYING 7:21
CYCLE COMPLETE 9:42

Digital Signature #
8950AEA773567473
