
Midmark M11 - v1.0.5
Total Cycles: 2828

Name: RITTER2
Sterilizer ID: V2313228
09/05/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.2 F 0.0
2:00 198.0 F 2.2
4:00 192.5 F 6.9
6:00 206.3 F 12.5
8:00 221.5 F 7.3
10:00 236.0 F 12.9
12:00 237.9 F 10.7
14:00 236.5 F 9.4
16:00 250.5 F 16.1
18:00 263.0 F 23.3
20:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.6
2:00 271.4 F 29.1
3:00 272.5 F 29.6
4:00 271.4 F 28.9

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 214.8 F 0.0
10:00 221.7 F 0.0
15:00 219.6 F 0.0
20:00 217.7 F 0.0
25:00 218.0 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

09/05/2023 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____