
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000351
13:42 19/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:02
STERILIZING 3:56
136.0 C 220kPa 3:56
136.4 C 223kPa 4:26
136.4 C 226kPa 4:56
136.4 C 224kPa 5:26
136.0 C 223kPa 5:56
136.4 C 222kPa 6:26
136.4 C 223kPa 6:56
136.4 C 225kPa 7:26
136.4 C 225kPa 7:26
VENTING 7:26
AIR DRYING 8:15
CYCLE COMPLETE 8:59

Digital Signature #
9CD9405F8CD59132
