

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001427  
13:05 09/13/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:51  
PRESSURIZING 5:57  
STERILIZING 7:34  
136.5 C 222kPa 7:34  
136.1 C 220kPa 8:04  
136.3 C 222kPa 8:34  
136.3 C 221kPa 9:04  
136.1 C 219kPa 9:34  
136.4 C 223kPa 10:04  
136.2 C 221kPa 10:34  
136.2 C 221kPa 11:04  
VENTING 11:04  
AIR DRYING 12:07  
CYCLE COMPLETE 42:07

-----  
Digital Signature #  
F8E7212C0905DB55  
-----