
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000895
15:26 26/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:43
STERILIZING 3:40
136.4 C 223kPa 3:40
136.1 C 219kPa 4:10
136.3 C 221kPa 4:40
136.4 C 224kPa 5:10
136.4 C 221kPa 5:40
136.4 C 223kPa 6:10
136.4 C 222kPa 6:40
136.1 C 222kPa 7:10
136.1 C 222kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 10:52

Digital Signature #
379B6BBF9B6C446D
