

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1127

Name: RITTER 1  
Sterilizer ID: V2178198  
08/30/2021 06:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.0 F 0.0  
2:00 73.8 F 0.0  
4:00 78.1 F 0.0  
6:00 114.9 F 0.1  
8:00 149.0 F 0.1  
10:00 168.6 F 0.2  
12:00 196.6 F 1.1  
14:00 215.2 F 4.3  
16:00 219.4 F 6.8  
18:00 228.1 F 10.0  
20:00 235.3 F 10.3  
22:00 238.5 F 10.6  
24:00 234.5 F 8.0  
26:00 241.1 F 10.5  
28:00 247.9 F 14.0  
30:00 260.4 F 20.8  
32:00 271.3 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.6 F 28.6  
2:00 271.6 F 28.2  
3:00 272.6 F 28.6  
4:00 271.5 F 28.0  
  
Min 271.4 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
5:00 185.4 F 0.0  
10:00 182.8 F 0.0  
15:00 184.3 F 0.0  
20:00 183.5 F 0.0  
25:00 181.8 F 0.0  
30:00 180.1 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:40  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:10:33  
CYCLE COMPLETE

08/30/2021 08:24 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02