

-----  
Midmark M11 - v1.0.5  
Total Cycles: 652

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2021 11:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.2 F 0.0  
2:00 171.5 F 3.6  
4:00 192.7 F 8.1  
6:00 211.6 F 13.2  
8:00 223.4 F 7.5  
10:00 236.3 F 12.8  
12:00 237.9 F 10.7  
14:00 237.3 F 9.6  
16:00 251.3 F 16.4  
18:00 263.0 F 23.2  
20:00 272.5 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.7 F 29.6  
2:00 271.7 F 28.9  
3:00 272.4 F 29.8  
4:00 272.5 F 29.4

Min 271.5 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
5:00 233.4 F 0.0  
10:00 225.9 F 0.0  
15:00 223.4 F 0.0  
20:00 215.7 F 0.0  
25:00 212.4 F 0.0  
30:00 209.4 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 20:14  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:02  
CYCLE COMPLETE

12/20/2021 12:44 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02