
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002050
8:15 18/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:03
STERILIZING 3:15
136.4 C 224kPa 3:15
136.4 C 225kPa 3:45
136.1 C 221kPa 4:15
136.4 C 223kPa 4:45
136.3 C 224kPa 5:15
136.4 C 225kPa 5:45
136.3 C 225kPa 6:15
136.4 C 225kPa 6:45
136.4 C 225kPa 6:45
VENTING 6:45
AIR DRYING 7:33
CYCLE COMPLETE 22:36

Digital Signature #
EFDBAB99F7F028AE
