
Midmark M11 - v1.0.0
Total Cycles: 2672

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2022 08:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.3 F 0.0
2:00 79.0 F 0.0
4:00 87.5 F 0.0
6:00 118.8 F 1.1
8:00 145.3 F 2.5
10:00 173.2 F 4.8
12:00 194.4 F 8.4
14:00 208.2 F 12.3
16:00 215.2 F 5.0
18:00 227.5 F 9.1
20:00 238.0 F 13.5
22:00 238.0 F 10.1
24:00 236.8 F 8.5
26:00 248.2 F 14.0
28:00 259.5 F 20.3
30:00 269.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.7
2:00 271.3 F 28.0
3:00 272.5 F 28.8
4:00 271.6 F 28.1

Min 271.2 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 204.2 F 0.0
10:00 222.8 F 0.0
15:00 223.9 F 0.0
20:00 224.5 F 0.0
25:00 231.3 F 0.0
30:00 218.2 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

12/14/2022 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____