
Midmark M11 - v1.0.0
Total Cycles: 4056

Name: RITTER 1
Sterilizer ID: V2178198
11/24/2023 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 72.8 F 0.0
4:00 108.1 F 0.4
6:00 135.0 F 3.2
8:00 145.8 F 6.8
10:00 171.7 F 10.9
12:00 195.8 F 2.1
14:00 211.8 F 5.2
16:00 223.7 F 8.9
18:00 237.5 F 13.0
20:00 234.6 F 8.8
22:00 241.8 F 11.0
24:00 244.9 F 12.4
26:00 257.8 F 19.2
28:00 269.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.1 F 28.1
2:00 272.7 F 28.7
3:00 271.5 F 27.9
4:00 272.4 F 28.3

Min 271.3 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 209.9 F 0.0
10:00 214.4 F 0.0
15:00 219.2 F 0.0
20:00 218.2 F 0.0
25:00 218.9 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 28:58
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:06:55
CYCLE COMPLETE

11/24/2023 12:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____