
Midmark M11 - v1.0.5
Total Cycles: 3607

Name: RITTER2
Sterilizer ID: V2313228
05/10/2024 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.5 F 0.0
2:00 111.8 F 0.7
4:00 131.0 F 3.0
6:00 159.8 F 5.9
8:00 183.2 F 9.3
10:00 198.5 F 12.9
12:00 208.1 F 5.3
14:00 219.0 F 9.0
16:00 231.2 F 13.1
18:00 232.9 F 9.5
20:00 238.5 F 10.7
22:00 243.4 F 12.8
24:00 254.2 F 18.4
26:00 264.2 F 24.4
28:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.5 F 30.1
2:00 271.6 F 29.4
3:00 272.0 F 30.1
4:00 272.3 F 29.9

Min 271.3 F 29.3
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 223.5 F 0.0
10:00 226.6 F 0.0
15:00 223.3 F 0.0
20:00 228.4 F 0.0
25:00 232.8 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 28:14
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:06:57
CYCLE COMPLETE

05/10/2024 03:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____