

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3511

Name: RITTER2  
Sterilizer ID: V2313228  
04/10/2024 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 78.6 F 0.0  
4:00 124.8 F 0.0  
6:00 106.7 F 0.2  
8:00 141.2 F 2.4  
10:00 173.7 F 5.6  
12:00 202.5 F 9.4  
14:00 210.6 F 8.7  
16:00 223.4 F 7.8  
18:00 233.6 F 12.2  
20:00 234.1 F 8.7  
22:00 240.5 F 11.1  
24:00 244.7 F 12.9  
26:00 257.0 F 19.5  
28:00 268.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 29.5  
2:00 272.6 F 29.9  
3:00 272.1 F 29.2  
4:00 273.2 F 30.0

Min 271.5 F 29.1  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.9 F 0.0  
5:00 220.0 F 0.0  
10:00 199.1 F 0.0  
15:00 196.0 F 0.0  
20:00 192.7 F 0.0  
25:00 190.7 F 0.0  
30:00 189.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:58  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:08:01  
CYCLE COMPLETE

04/10/2024 09:52 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_