
Midmark M11 - v1.0.0
Total Cycles: 2277

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2022 05:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 171.4 F 0.5
4:00 188.4 F 3.2
6:00 210.2 F 7.8
8:00 225.4 F 13.1
10:00 230.2 F 8.6
12:00 239.6 F 11.5
14:00 242.3 F 12.0
16:00 240.7 F 10.8
18:00 254.6 F 17.8
20:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.0
2:00 271.5 F 28.3
3:00 272.7 F 29.1
4:00 271.7 F 28.3

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 216.1 F 0.0
10:00 234.0 F 0.0
15:00 229.9 F 0.0
20:00 229.3 F 0.0
25:00 228.6 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

08/31/2022 05:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____