
Midmark M11 - v1.0.0
Total Cycles: 5450

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2025 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.5 F 0.0
2:00 139.8 F 2.1
4:00 155.2 F 5.6
6:00 179.1 F 9.5
8:00 197.8 F 9.2
10:00 213.9 F 6.0
12:00 227.9 F 10.4
14:00 232.0 F 7.4
16:00 240.3 F 11.6
18:00 237.7 F 9.7
20:00 250.6 F 15.9
22:00 261.7 F 22.3
24:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.2 F 29.4
2:00 272.3 F 29.0
3:00 271.3 F 28.5
4:00 271.5 F 29.1

Min 271.1 F 28.5
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 206.5 F 0.0
10:00 206.8 F 0.0
15:00 202.7 F 0.0
20:00 210.2 F 0.0
25:00 209.1 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:03:31
CYCLE COMPLETE

03/26/2025 03:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____