
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000697
9:48 16/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:43
STERILIZING 3:45
136.5 C 232kPa 3:45
136.3 C 232kPa 4:15
136.0 C 232kPa 4:45
136.4 C 233kPa 5:15
136.4 C 232kPa 5:45
136.2 C 233kPa 6:15
136.4 C 230kPa 6:45
136.3 C 233kPa 7:15
136.3 C 233kPa 7:15
VENTING 7:15
AIR DRYING 8:08
CYCLE COMPLETE 8:14

Digital Signature #
2F8EC8291E81247C
