
Midmark M11 - v1.0.5
Total Cycles: 2902

Name: RITTER2
Sterilizer ID: V2313228
09/18/2023 04:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.6 F 0.0
2:00 167.1 F 3.6
4:00 194.4 F 8.7
6:00 212.6 F 8.1
8:00 226.2 F 10.1
10:00 232.2 F 8.2
12:00 240.3 F 11.3
14:00 246.4 F 14.1
16:00 259.7 F 21.2
18:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.1 F 29.4
2:00 271.8 F 29.6
3:00 271.5 F 29.1
4:00 272.2 F 29.8

Min 271.1 F 28.8
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 220.5 F 0.0
10:00 223.2 F 0.0
15:00 223.3 F 0.0
20:00 220.7 F 0.0
25:00 219.1 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:40
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

09/18/2023 05:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____