
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008010
8:45 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:59
PRESSURIZING 3:59
STERILIZING 4:56
136.1 C 220kPa 4:56
136.1 C 221kPa 5:26
136.3 C 222kPa 5:56
136.4 C 223kPa 6:26
136.1 C 221kPa 6:56
136.2 C 222kPa 7:26
136.1 C 221kPa 7:56
136.4 C 221kPa 8:26
136.4 C 221kPa 8:26
VENTING 8:26
AIR DRYING 9:15
CYCLE COMPLETE 39:15

Digital Signature #
AB63BCA5543AC55C
