

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2704

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2023 10:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.7 F 0.0  
2:00 171.1 F 2.2  
4:00 181.5 F 6.0  
6:00 200.9 F 10.3  
8:00 215.7 F 5.6  
10:00 225.2 F 8.9  
12:00 236.7 F 12.3  
14:00 238.9 F 11.2  
16:00 237.7 F 9.8  
18:00 250.9 F 16.2  
20:00 262.1 F 22.7  
22:00 272.3 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.7  
2:00 271.9 F 29.0  
3:00 271.7 F 29.4  
4:00 272.4 F 29.6

Min 271.3 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
5:00 197.4 F 0.0  
10:00 202.8 F 0.0  
15:00 203.2 F 0.0  
20:00 200.5 F 0.0  
25:00 195.4 F 0.0  
30:00 192.4 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 22:29  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:01:14  
CYCLE COMPLETE

07/26/2023 11:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_