
Midmark M11 - v1.0.5
Total Cycles: 989

Name: RITTER2
Sterilizer ID: V2313228
04/14/2022 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 150.9 F 0.1
4:00 165.1 F 0.6
6:00 197.8 F 4.3
8:00 213.3 F 9.5
10:00 229.5 F 8.8
12:00 235.8 F 11.4
14:00 236.4 F 9.8
16:00 237.3 F 9.0
18:00 251.2 F 16.4
20:00 263.7 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.5
2:00 271.4 F 29.1
3:00 272.4 F 29.6
4:00 271.5 F 28.9

Min 271.4 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 193.1 F 0.0
10:00 193.6 F 0.0
15:00 194.0 F 0.0
20:00 196.0 F 0.0
25:00 195.9 F 0.0
30:00 196.0 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

04/14/2022 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____