
Midmark M11 - v1.0.0
Total Cycles: 2724

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2022 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.9 F 0.0
2:00 211.8 F 3.5
4:00 214.2 F 8.6
6:00 225.5 F 9.0
8:00 231.7 F 10.1
10:00 236.9 F 7.9
12:00 243.3 F 12.2
14:00 245.1 F 12.9
16:00 258.2 F 19.9
18:00 269.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.3 F 29.1
2:00 271.7 F 28.3
3:00 272.3 F 29.1
4:00 271.9 F 28.5

Min 271.2 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.0 F 0.0
5:00 230.0 F 0.0
10:00 238.9 F 0.0
15:00 232.1 F 0.0
20:00 227.3 F 0.0
25:00 228.0 F 0.0
30:00 228.4 F 0.0

Drying Complete

FILLING: 1:48
HEATING: 18:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:45
CYCLE COMPLETE

12/28/2022 12:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____