

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2338

Name: RITTER2  
Sterilizer ID: V2313228  
05/04/2023 10:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.0 F 0.0  
2:00 174.5 F 3.3  
4:00 186.9 F 6.9  
6:00 208.0 F 11.8  
8:00 213.3 F 6.1  
10:00 229.7 F 11.5  
12:00 231.7 F 8.4  
14:00 239.6 F 11.1  
16:00 244.5 F 13.3  
18:00 257.5 F 20.0  
20:00 268.3 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.3 F 30.0  
2:00 271.9 F 29.3  
3:00 271.2 F 29.3  
4:00 272.3 F 30.0  
  
Min 271.1 F 29.0  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.8 F 0.0  
5:00 211.8 F 0.0  
10:00 219.0 F 0.0  
15:00 219.2 F 0.0  
20:00 217.3 F 0.0  
25:00 218.3 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:14  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:36  
CYCLE COMPLETE

05/04/2023 11:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_