
Midmark M11 - v1.0.0
Total Cycles: 5343

Name: RITTER 1
Sterilizer ID: V2178198
02/19/2025 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.4 F 0.0
2:00 142.8 F 1.5
4:00 150.8 F 4.4
6:00 160.2 F 7.7
8:00 184.1 F 11.4
10:00 212.7 F 2.7
12:00 218.2 F 6.3
14:00 228.1 F 10.0
16:00 235.3 F 9.9
18:00 237.6 F 10.5
20:00 237.7 F 8.6
22:00 246.9 F 13.7
24:00 257.3 F 19.3
26:00 267.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.1 F 29.1
2:00 271.9 F 28.6
3:00 271.0 F 28.5
4:00 272.3 F 29.6

Min 270.9 F 28.3
Max 273.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 212.4 F 0.0
10:00 212.7 F 0.0
15:00 214.8 F 0.0
20:00 220.8 F 0.0
25:00 222.3 F 0.0
30:00 229.7 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 27:36
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:06:28
CYCLE COMPLETE

02/19/2025 02:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____