
Midmark M11 - v1.0.5
Total Cycles: 3640

Name: RITTER2
Sterilizer ID: V2313228
05/23/2024 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.1 F 0.0
2:00 99.2 F 0.2
4:00 115.2 F 1.7
6:00 140.3 F 4.6
8:00 170.8 F 8.0
10:00 191.6 F 11.8
12:00 208.8 F 3.5
14:00 219.1 F 6.8
16:00 230.8 F 10.8
18:00 231.7 F 7.8
20:00 242.7 F 12.8
22:00 241.6 F 11.5
24:00 253.5 F 17.6
26:00 264.1 F 23.9
28:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.6
2:00 271.8 F 28.9
3:00 272.1 F 29.6
4:00 272.4 F 29.4

Min 271.5 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.0 F 0.0
5:00 217.5 F 0.0
10:00 204.1 F 0.0
15:00 191.9 F 0.0
20:00 188.3 F 0.0
25:00 186.8 F 0.0
30:00 199.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:10
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:06:38
CYCLE COMPLETE

05/23/2024 04:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____