
Midmark M11 - v1.0.5
Total Cycles: 3487

Name: RITTER2
Sterilizer ID: V2313228
03/28/2024 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.0 F 0.0
2:00 89.7 F 0.0
4:00 114.9 F 0.1
6:00 148.2 F 1.2
8:00 169.9 F 4.3
10:00 190.0 F 7.9
12:00 204.2 F 11.7
14:00 216.4 F 4.9
16:00 226.8 F 8.8
18:00 237.5 F 13.3
20:00 238.2 F 10.6
22:00 235.9 F 8.9
24:00 249.6 F 15.4
26:00 260.9 F 21.8
28:00 271.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.8 F 29.4
2:00 271.5 F 28.8
3:00 272.9 F 29.5
4:00 271.9 F 28.8

Min 271.5 F 28.6
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.9 F 0.0
5:00 205.3 F 0.0
10:00 212.0 F 0.0
15:00 205.5 F 0.0
20:00 200.3 F 0.0
25:00 196.8 F 0.0
30:00 193.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:40
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:06:57
CYCLE COMPLETE

03/28/2024 01:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____