
Midmark M11 - v1.0.5
Total Cycles: 1733

Name: RITTER2
Sterilizer ID: V2313228
11/16/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.6 F 0.0
2:00 194.1 F 3.1
4:00 206.3 F 8.6
6:00 221.7 F 9.7
8:00 231.5 F 10.6
10:00 233.0 F 8.7
12:00 238.6 F 10.0
14:00 248.5 F 15.1
16:00 261.3 F 22.3
18:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.8 F 29.9
2:00 272.0 F 29.1
3:00 271.9 F 29.7
4:00 272.6 F 29.8

Min 271.4 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 224.2 F 0.0
10:00 244.8 F 0.0
15:00 229.2 F 0.0
20:00 222.6 F 0.0
25:00 221.6 F 0.0
30:00 215.2 F 0.0

Drying Complete

FILLING: 1:50
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:26
CYCLE COMPLETE

11/16/2022 01:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____