
Midmark M11 - v1.0.5
Total Cycles: 2195

Name: RITTER2
Sterilizer ID: V2313228
03/30/2023 07:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 80.6 F 0.0
4:00 85.8 F 0.0
6:00 111.2 F 0.0
8:00 150.9 F 2.7
10:00 175.8 F 5.8
12:00 201.0 F 9.3
14:00 209.0 F 11.1
16:00 221.4 F 7.1
18:00 233.5 F 11.4
20:00 231.3 F 8.0
22:00 243.0 F 12.4
24:00 243.4 F 12.4
26:00 255.2 F 18.6
28:00 265.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.4 F 29.8
2:00 271.5 F 29.0
3:00 272.0 F 29.7
4:00 272.3 F 29.5

Min 271.1 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 206.9 F 0.0
10:00 220.0 F 0.0
15:00 196.0 F 0.0
20:00 191.8 F 0.0
25:00 189.2 F 0.0
30:00 192.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:44
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:07:45
CYCLE COMPLETE

03/30/2023 08:26 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____