
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.4uS / 1.5ppm
CYCLE NUMBER 000015
15:05 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:55
PRESSURIZING 6:08
STERILIZING 7:51
136.0 C 223kPa 7:51
136.5 C 227kPa 8:21
136.5 C 227kPa 8:52
136.4 C 228kPa 9:22
136.4 C 226kPa 9:52
136.4 C 227kPa 10:22
136.4 C 228kPa 10:52
136.2 C 227kPa 11:21
VENTING 11:22
AIR DRYING 12:37
CYCLE COMPLETE 14:37

Digital Signature #
94C6BC3FD167EEC6
