
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001776
7:27 16/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:57
STERILIZING 4:06
136.5 C 222kPa 4:06
136.4 C 223kPa 4:36
136.3 C 223kPa 5:06
136.4 C 221kPa 5:36
136.4 C 219kPa 6:06
136.4 C 221kPa 6:36
136.2 C 223kPa 7:06
136.4 C 223kPa 7:36
136.4 C 223kPa 7:36
VENTING 7:36
AIR DRYING 8:28
CYCLE COMPLETE 8:38

Digital Signature #
B481046BE2AE02E2
