
Midmark M11 - v1.0.5
Total Cycles: 430

Name: RITTER2
Sterilizer ID: V2313228
10/08/2021 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 73.8 F 0.0
4:00 76.7 F 0.0
6:00 91.7 F 0.1
8:00 147.1 F 0.2
10:00 166.8 F 3.4
12:00 183.9 F 7.0
14:00 203.2 F 10.7
16:00 221.0 F 7.5
18:00 229.1 F 9.2
20:00 238.0 F 12.4
22:00 239.6 F 11.1
24:00 237.4 F 9.4
26:00 250.4 F 15.7
28:00 260.7 F 21.8
30:00 270.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 271.6 F 29.4
2:00 271.9 F 29.4
3:00 272.1 F 29.2
4:00 271.9 F 29.0

Min 271.0 F 28.9
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 218.2 F 0.0
10:00 209.1 F 0.0
15:00 198.8 F 0.0
20:00 205.8 F 0.0
25:00 221.5 F 0.0
30:00 204.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 30:38
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:40
CYCLE COMPLETE

10/08/2021 07:52 AM
ACCEPTED: X
REJECTED:
Operator: U02