
Midmark M11 - v1.0.5
Total Cycles: 3086

Name: RITTER2
Sterilizer ID: V2313228
11/06/2023 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 160.2 F 1.1
4:00 169.1 F 4.2
6:00 189.4 F 8.3
8:00 209.4 F 13.0
10:00 218.3 F 7.1
12:00 232.0 F 11.8
14:00 235.1 F 8.8
16:00 239.8 F 10.8
18:00 245.5 F 13.2
20:00 257.0 F 19.4
22:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.3 F 29.5
2:00 271.1 F 28.8
3:00 272.4 F 29.4
4:00 271.0 F 28.8

Min 271.0 F 28.7
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 202.3 F 0.0
10:00 200.7 F 0.0
15:00 197.7 F 0.0
20:00 194.1 F 0.0
25:00 191.2 F 0.0
30:00 189.1 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

11/06/2023 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____